



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client-site
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(50 to 500) mbar	0.004 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(100 to 1000) mbar	0.004 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(>1 to 20) bar	0.008 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(20 to 210) bar	0.03 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(-0.9 to -0.1) bar	0.004 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(0 to 1000) mbar	0.004 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(>1 to 20) bar	0.009 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(20 to 210) bar	0.03 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Pressure	Hydraulic Pressure Gauge/ Differential Pressure Gauge	(0 to 35) bar	0.02 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Hydraulic Pressure Gauge/ Differential Pressure Gauge	(35 to 100) bar	0.04 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Hydraulic Pressure Gauge/ Differential Pressure Gauge	(100 to 700) bar	0.2 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Hydraulic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(0 to 35) bar	0.02 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Pressure	Hydraulic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(35 to 100) bar	0.04 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Hydraulic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(100 to 700) bar	0.2 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Vacuum Gauge (Analog & Digital)	(0 to -0.9) bar	0.004 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with RTD	(-40 to 100) °C	0.11 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with RTD	(>100 to 400) °C	0.21 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with RTD	(>400 to 650) °C	0.63 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Thermodynamics and temperature measurements	Temperature Transmitter with RTD	(-40 to 100) °C	0.11 °C	API-QLT-WI-04 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Transmitter with RTD	(>100 to 400) °C	0.21 °C	API-QLT-WI-04 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Transmitter with RTD	(>400 to 650) °C	0.63 °C	API-QLT-WI-04 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Transmitter with TC	(-40 to 100) °C	0.24 °C	API-QLT-WI-04 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Transmitter with TC	(>100 to 400) °C	0.82 °C	API-QLT-WI-04 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Transmitter with TC	(>400 to 650) °C	1.4 °C	API-QLT-WI-04 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with TC	(-40 to 100) °C	0.24 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with TC	(>100 to 400) °C	0.82 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with TC	(>400 to 650) °C	1.4 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Thermodynamics and temperature measurements	Dial Thermometer	(-40 to 100) °C	0.39 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Dial Thermometer	(>100 to 400) °C	1.1 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Dial Thermometer	(>400 to 650) °C	2.0 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
END					

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.