



# Accreditation Scope

VTIC ONSHORE & OFFSHORE OIL & GAS SERVICES, NAL 226  
Testing Laboratory, (ISO/IEC 17025:2017)

Warehouse 117 A, P.O Box 50244, Fujairah Freezone Phase 1, Fujairah, UAE. , Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

Testing Field	Materials/ Products Tested	Type of test/ Test parameter/ Properties measured/Range of measurement	Test Method (Standard, Internal Procedure, Technique)	Permanent lab (P) / Client-site (S)
Petroleum and related technologies	Fuel Oil / Gas Oil	Density @15° C (Digital Density Meter) 0.6000 - 1.1000 Kg/L	ISO 12185-2024 / ASTM D 4052-2022	P
Petroleum and related technologies	Fuel Oil / Gas Oil	Kinematic viscosity at various temperature (0.2 to 300000mm <sup>2</sup> /sec)	ISO 3104-2023/ ASTM D 445-2024	P
Petroleum and related technologies	Fuel Oil / Gas Oil	Flash Point PMCC (40 to 360 degree celcius)	ISO 2719-2016 A/ ASTM D93-2020 Methods A and B	P
Petroleum and related technologies	Fuel Oil / Gas Oil	Sulphur (17 ppm to 4.6%m)	ISO 8754-2003/ ASTM D 4294-2021	P
Petroleum and related technologies	Fuel Oil / Gas Oil	Water by Distillation (0 to 25 %v)	ISO 3733-1999/ ASTM D 95-2023	P
Petroleum and related technologies	Fuel Oil / Gas Oil	Pour Point (minus 38 to plus 50 degree celcius)	ISO 3016-2019 / ASTM D 97-2017 (2022)	P
Petroleum and related technologies	Fuel Oil / Gas Oil	Hydrogen Sulphide (0.40 ppm to 15 ppm)	IP 570-2021 / ASTM D 7621-2021 Method A	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



# Accreditation Scope

VTIC ONSHORE & OFFSHORE OIL & GAS SERVICES, NAL 226  
Testing Laboratory, (ISO/IEC 17025:2017)

Warehouse 117 A, P.O Box 50244, Fujairah Freezone Phase 1, Fujairah, UAE. , Fujairah, United Arab Emirates

Issue Date: 23/12/2024

Expiry Date: 22/12/2027

Issue No: 1

END

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.