



Accreditation Scope

FANR Secondary Standards Dosimetry Laboratory, NAL131
Calibration Laboratory, (ISO/IEC 17025:2017)

Abu Dhabi, Al Zafranah area next to Khalifa University, Abu Dhabi, United Arab Emirates

Issue Date: 04/06/2025

Expiry Date: 05/12/2025

Issue No: 5

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client- site
Ionizing Radiation measurements	Air Kerma Rate / Radioprotection Dosemeter	(1.0E-6 – 0.04) Gy/h	1.4 - 2.4 %	ISO 4037:2019 with Cs-137 source	P
Ionizing Radiation measurements	Ambient Dose Equivalent Rate / Active Dosemeters	(2 – 50,000) μSv/h	4.7 - 5.2 %	ISO 4037:2019 with Cs-137 source, Free in Air	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 10 mm depth) with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.3 %	ISO 4037:2019 with Cs-137 source, ISO slab phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 0.07 mm depth) with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.3 %	ISO 4037:2019 with Cs-137 source, ISO Rod or Pillar or Slab phantom	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

FANR Secondary Standards Dosimetry Laboratory, NAL131
Calibration Laboratory, (ISO/IEC 17025:2017)

Abu Dhabi, Al Zafranah area next to Khalifa University, Abu Dhabi, United Arab Emirates

Issue Date: 04/06/2025

Expiry Date: 05/12/2025

Issue No: 5

Ionizing Radiation measurements	Personal Dose Equivalent (in 3 mm depth) with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.3 %	ISO 4037:2019 with Cs-137 source, ISO cylinder phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 10mm depth) with angular dependence / Active Dosimeters	(0.01 to 1,000) mSv	4.7 - 6.6 %	ISO 4037:2019 with Cs-137 source, ISO slab phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent Rate (in 10mm depth) with angular dependence / Active Dosimeters	(2 – 50,000) μ Sv/h	4.7 - 6.6 %	ISO 4037:2019 with Cs-137 source, ISO slab phantom	P
Ionizing Radiation measurements	Ambient Dose Equivalent with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.3 %	ISO 4037:2019 with Cs-137 source, Free in Air	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

FANR Secondary Standards Dosimetry Laboratory, NAL131
Calibration Laboratory, (ISO/IEC 17025:2017)

Abu Dhabi, Al Zafranah area next to Khalifa University, Abu Dhabi, United Arab Emirates

Issue Date: 04/06/2025

Expiry Date: 05/12/2025

Issue No: 5

Ionizing Radiation measurements	Air Kerma Rate / Diagnostic Dosimeters	(1.3 – 3.3) Gy/h	1.9 %	IEC 61267:2005, X-Ray RQR Qualities, 50 kV to 150 kV	P
Ionizing Radiation measurements	Air Kerma Rate / Diagnostic Dosimeters	(0.2 – 0.5) Gy/h	1.9 %	IEC 61267:2005, X-Ray RQA Qualities, 50 kV to 150 kV	P
Ionizing Radiation measurements	Air Kerma Rate / Radioprotection Dosimeter	(3.6E-04 – 2.3E-02) Gy/h	1.9 %	ISO 4037:2019 , X-ray N-Series, 40 kV to 300 kV	P
Ionizing Radiation measurements	Ambient Dose Equivalent Rate / Active Dosimeters	(200 – 10,000) μ Sv/h	4.5 - 4.9 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, Free in Air	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 10 mm depth) with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.0 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, ISO slab phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 0.07 mm depth) with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.0 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, ISO pillar, rod and slab phantoms	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

FANR Secondary Standards Dosimetry Laboratory, NAL131
Calibration Laboratory, (ISO/IEC 17025:2017)

Abu Dhabi, Al Zafranah area next to Khalifa University, Abu Dhabi, United Arab Emirates

Issue Date: 04/06/2025

Expiry Date: 05/12/2025

Issue No: 5

Ionizing Radiation measurements	Personal Dose Equivalent (in 10mm depth) with angular dependence / Active Dosimeters	(0.01 to 1,000) mSv	4.5 - 6.3 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, ISO slab phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent Rate (in 10mm depth) with angular dependence / Active Dosimeters	(100 – 10,000) μ Sv/h	4.5 - 6.3 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, ISO slab phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 3 mm depth) with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.0 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, ISO cylinder phantom	P
Ionizing Radiation measurements	Ambient Dose Equivalent with angular dependence / Passive Dosimeters	(0.01 to 1,000) mSv	6.0 %	ISO 4037:2019, X-ray N-Series, 40 kV to 300 kV, Free in Air	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

FANR Secondary Standards Dosimetry Laboratory, NAL131
Calibration Laboratory, (ISO/IEC 17025:2017)

Abu Dhabi, Al Zafranah area next to Khalifa University, Abu Dhabi, United Arab Emirates

Issue Date: 04/06/2025

Expiry Date: 05/12/2025

Issue No: 5

Ionizing Radiation measurements	Ambient Dose Equivalent Rate / Active Dosimeters (Neutron Measurements)	(20.0 - 2,323) μ Sv/h	8.7 %	ISO 8529-1:2021 / 2:2000 / 3:2023, AmBe-241, Free in Air	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 10 mm depth) with angular dependence / Active Dosimeters (Neutron Measurements)	(0.3 to 30) mSv	9.9 %	ISO 8529-1:2021 / 2:2000 / 3:2023, AmBe-241, ISO slab phantom	P
Ionizing Radiation measurements	Personal Dose Equivalent (in 10 mm depth) with angular dependence / Passive Dosimeters (Neutron Measurements)	(0.3 to 30) mSv	9.9 %	ISO 8529-1:2021 / 2:2000 / 3:2023, AmBe-241, ISO slab phantom	P
END					

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.